



	1	2	3	4	5	6	7	8	9	10	11	12
F	Unk FAM	Unk FAM gDNA	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM
G	Unk FAM	NTC FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM
H	Unk FAM	NTC FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM	Unk FAM

## Quantification

Step #: 3

Analysis Mode: Fluorophore


Cq Determination: Single Threshold

Baseline Method:

FAM: Auto Calculated

Threshold Setting:

FAM: None, Auto Calculated



## Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
A01	FAM		Unkn		N/A	0.00	0.000
A02	FAM		Unkn		N/A	0.00	0.000
A03	FAM		NTC		N/A	0.00	0.000
A04	FAM		Unkn		N/A	0.00	0.000
A05	FAM		Unkn		N/A	0.00	0.000
A06	FAM		Unkn		N/A	0.00	0.000
A07	FAM		Unkn		N/A	0.00	0.000
A08	FAM		Unkn		N/A	0.00	0.000

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
A09	FAM		Unkn		N/A	0.00	0.000
A10	FAM		Unkn		N/A	0.00	0.000
A11	FAM		Unkn		N/A	0.00	0.000
A12	FAM		Unkn		N/A	0.00	0.000
B01	FAM		Unkn		N/A	0.00	0.000
B02	FAM		Unkn		N/A	0.00	0.000
B03	FAM		Unkn		N/A	0.00	0.000
B04	FAM		Unkn		N/A	0.00	0.000
B05	FAM		Unkn		N/A	0.00	0.000
B06	FAM		Unkn		N/A	0.00	0.000
B07	FAM		Unkn		N/A	0.00	0.000
B08	FAM		Unkn		N/A	0.00	0.000
B09	FAM		Unkn		N/A	0.00	0.000
B10	FAM		Unkn		N/A	0.00	0.000
B11	FAM		Unkn		N/A	0.00	0.000
B12	FAM		Unkn		N/A	0.00	0.000
C01	FAM		Unkn		N/A	0.00	0.000
C02	FAM		Unkn		N/A	0.00	0.000
C03	FAM		Unkn		N/A	0.00	0.000
C04	FAM		Unkn		N/A	0.00	0.000
C05	FAM		Unkn		N/A	0.00	0.000
C06	FAM		Unkn		N/A	0.00	0.000
C07	FAM		Unkn		N/A	0.00	0.000
C08	FAM		Unkn		N/A	0.00	0.000
C09	FAM		Unkn		N/A	0.00	0.000
C10	FAM		Unkn		N/A	0.00	0.000
C11	FAM		Unkn		N/A	0.00	0.000
C12	FAM		Unkn		N/A	0.00	0.000
D01	FAM		Unkn		N/A	0.00	0.000
D02	FAM		Unkn		N/A	0.00	0.000
D03	FAM		Unkn		N/A	0.00	0.000
D04	FAM		Unkn		N/A	0.00	0.000
D05	FAM		Unkn		N/A	0.00	0.000
D06	FAM		Unkn		N/A	0.00	0.000
D07	FAM		Unkn		N/A	0.00	0.000
D08	FAM		Unkn		N/A	0.00	0.000
D09	FAM		Unkn		N/A	0.00	0.000
D10	FAM		Unkn		N/A	0.00	0.000
D11	FAM		Unkn		N/A	0.00	0.000
D12	FAM		Unkn		N/A	0.00	0.000
E01	FAM		Unkn		N/A	0.00	0.000
E02	FAM		Unkn	gDNA	N/A	0.00	0.000
E03	FAM		Unkn		N/A	0.00	0.000
E04	FAM		Unkn		N/A	0.00	0.000
E05	FAM		Unkn		N/A	0.00	0.000
E06	FAM		Unkn		N/A	0.00	0.000
E07	FAM		Unkn		N/A	0.00	0.000
E08	FAM		Unkn		N/A	0.00	0.000
E09	FAM		Unkn		N/A	0.00	0.000


Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
E10	FAM		Unkn		N/A	0.00	0.000
E11	FAM		Unkn		N/A	0.00	0.000
E12	FAM		Unkn		N/A	0.00	0.000
F01	FAM		Unkn		N/A	0.00	0.000
F02	FAM		Unkn	gDNA	N/A	0.00	0.000
F03	FAM		Unkn		N/A	0.00	0.000
F04	FAM		Unkn		N/A	0.00	0.000
F05	FAM		Unkn		N/A	0.00	0.000
F06	FAM		Unkn		N/A	0.00	0.000
F07	FAM		Unkn		N/A	0.00	0.000
F08	FAM		Unkn		N/A	0.00	0.000
F09	FAM		Unkn		N/A	0.00	0.000
F10	FAM		Unkn		N/A	0.00	0.000
F11	FAM		Unkn		N/A	0.00	0.000
F12	FAM		Unkn		N/A	0.00	0.000
G01	FAM		Unkn		N/A	0.00	0.000
G02	FAM		NTC		N/A	0.00	0.000
G03	FAM		Unkn		N/A	0.00	0.000
G04	FAM		Unkn		N/A	0.00	0.000
G05	FAM		Unkn		N/A	0.00	0.000
G06	FAM		Unkn		N/A	0.00	0.000
G07	FAM		Unkn		N/A	0.00	0.000
G08	FAM		Unkn		N/A	0.00	0.000
G09	FAM		Unkn		N/A	0.00	0.000
G10	FAM		Unkn		N/A	0.00	0.000
G11	FAM		Unkn		N/A	0.00	0.000
G12	FAM		Unkn		N/A	0.00	0.000
H01	FAM		Unkn		N/A	0.00	0.000
H02	FAM		NTC		N/A	0.00	0.000
H03	FAM		Unkn		N/A	0.00	0.000
H04	FAM		Unkn		N/A	0.00	0.000
H05	FAM		Unkn		N/A	0.00	0.000
H06	FAM		Unkn		N/A	0.00	0.000
H07	FAM		Unkn		N/A	0.00	0.000
H08	FAM		Unkn		N/A	0.00	0.000
H09	FAM		Unkn		N/A	0.00	0.000
H10	FAM		Unkn		N/A	0.00	0.000
H11	FAM		Unkn		N/A	0.00	0.000
H12	FAM		Unkn		N/A	0.00	0.000


## Gene Expression

Expression analysis is not possible, wells must contain at least two samples with target(s), which have a valid Cq.

## Melt Curve

Step #: 5





Melt Curve Data

Well	Fluor	Target	Content	Sample	Melt Temp
A01	FAM		Unkn		82.60
A01	FAM		Unkn		80.80
A02	FAM		Unkn		82.60
A02	FAM		Unkn		80.80
A03	FAM		NTC		None
A04	FAM		Unkn		None
A05	FAM		Unkn		None
A06	FAM		Unkn		None
A07	FAM		Unkn		None
A08	FAM		Unkn		None
A09	FAM		Unkn		None
A10	FAM		Unkn		None
A11	FAM		Unkn		None
A12	FAM		Unkn		None
B01	FAM		Unkn		82.60
B01	FAM		Unkn		80.80
B02	FAM		Unkn		82.60
B02	FAM		Unkn		81.00
B03	FAM		Unkn		None
B04	FAM		Unkn		None
B05	FAM		Unkn		None
B06	FAM		Unkn		None
B07	FAM		Unkn		None
B08	FAM		Unkn		None
B09	FAM		Unkn		None

Well	Fluor	Target	Content	Sample	Melt Temp
B10	FAM		Unkn		None
B11	FAM		Unkn		None
B12	FAM		Unkn		None
C01	FAM		Unkn		83.00
C02	FAM		Unkn		82.60
C02	FAM		Unkn		81.00
C03	FAM		Unkn		None
C04	FAM		Unkn		None
C05	FAM		Unkn		None
C06	FAM		Unkn		None
C07	FAM		Unkn		None
C08	FAM		Unkn		None
C09	FAM		Unkn		None
C10	FAM		Unkn		None
C11	FAM		Unkn		None
C12	FAM		Unkn		None
D01	FAM		Unkn		None
D02	FAM		Unkn		82.80
D02	FAM		Unkn		81.00
D03	FAM		Unkn		None
D04	FAM		Unkn		None
D05	FAM		Unkn		None
D06	FAM		Unkn		None
D07	FAM		Unkn		None
D08	FAM		Unkn		None
D09	FAM		Unkn		None
D10	FAM		Unkn		None
D11	FAM		Unkn		None
D12	FAM		Unkn		None
E01	FAM		Unkn		82.80
E01	FAM		Unkn		81.00
E02	FAM		Unkn	gDNA	None
E03	FAM		Unkn		None
E04	FAM		Unkn		None
E05	FAM		Unkn		None
E06	FAM		Unkn		None
E07	FAM		Unkn		None
E08	FAM		Unkn		None
E09	FAM		Unkn		None
E10	FAM		Unkn		None
E11	FAM		Unkn		None
E12	FAM		Unkn		None
F01	FAM		Unkn		82.80
F01	FAM		Unkn		81.00
F02	FAM		Unkn	gDNA	None
F03	FAM		Unkn		None
F04	FAM		Unkn		None
F05	FAM		Unkn		None
F06	FAM		Unkn		None
F07	FAM		Unkn		None

<b>Well</b>	<b>Fluor</b>	<b>Target</b>	<b>Content</b>	<b>Sample</b>	<b>Melt Temp</b>
F08	FAM		Unkn		None
F09	FAM		Unkn		None
F10	FAM		Unkn		None
F11	FAM		Unkn		None
F12	FAM		Unkn		None
G01	FAM		Unkn		82.80
G01	FAM		Unkn		80.80
G02	FAM		NTC		None
G03	FAM		Unkn		None
G04	FAM		Unkn		None
G05	FAM		Unkn		None
G06	FAM		Unkn		None
G07	FAM		Unkn		None
G08	FAM		Unkn		None
G09	FAM		Unkn		None
G10	FAM		Unkn		None
G11	FAM		Unkn		None
G12	FAM		Unkn		None
H01	FAM		Unkn		82.80
H01	FAM		Unkn		80.80
H02	FAM		NTC		None
H03	FAM		Unkn		None
H04	FAM		Unkn		None
H05	FAM		Unkn		None
H06	FAM		Unkn		None
H07	FAM		Unkn		None
H08	FAM		Unkn		None
H09	FAM		Unkn		None
H10	FAM		Unkn		None
H11	FAM		Unkn		None
H12	FAM		Unkn		None